



STATE PROCUREMENT OFFICE
NOTICE & REQUEST FOR SOLE SOURCE

14 FEB 11 AIO :54

STATE PROCUREMENT OFFICE
STATE OF HAWAII

TO: Chief Procurement Officer
FROM: Health / State Laboratories Division
Name of Requesting Department

Pursuant to HRS §103D-306 and HAR chapter 3-122, Subchapter 9, the Department requests sole source approval to purchase the following:

1. Describe the goods, services, or construction to be procured.
The PerkinElmer NexION 300 Inductively Coupled Plasma - Mass Spectrometer (ICP-MS) will replace the Chemical Response Laboratory's (CRL) nearly obsolete PerkinElmer ELAN DRC II ICP-MS. This system is used for measuring toxic elements in clinical, environmental and food samples. It is also part of a larger system that analyzes for mercury and arsenic speciation by Tandem HPLC-ICP-MS.

Table with 2 columns: Vendor/Contractor/Service Provider Name, Amount of Request, Term of contract, Prior SPO-001, Sole Source (SS) No. Includes handwritten dates for term of contract.

6. Describe in detail the following:
a. The unique features, characteristics, or capabilities of the goods, service or construction.
The PerkinElmer NexION 300 ICP-MS will replace and upgrade the current ICP-MS and will be compatible with the existing PerkinElmer tandem High Performance Liquid Chromatography - Inductively Coupled Plasma - Mass Spectrometer (HPLC-ICP-MS) system.
b. How the unique features, characteristics or capabilities of the goods, service or construction are essential for the department
The upgraded NexION 300 ICP-MS will replace the nearly obsolete ELAN DRC II ICP-MS which will not be supported by a service contract, and repair parts and supplies will be limited by 2017.

7. Describe the efforts and results in determining that this is the only vendor/contractor/service provider who can provide the goods, services or construction.

The PerkinElmer NexION 300 ICP-MS will replace and upgrade the current PerkinElmer ELAN DRC II ICP-MS. The NexION has become the primary platform for the three core CDC ICP-MS methods that the CRL is responsible for in order to meet program performance measures. CDC subject matter experts are only able to offer technical assistance for this particular system.

The CRL currently uses a PerkinElmer HPLC as part of its HPLC-ICP-MS system. The Chromera software that controls this HPLC is only compatible with and integrated into the PerkinElmer ELAN/NexION ICP-MS control software.

8. Alternate source. Describe the other possible sources for the goods, services, or construction that were investigated but did not meet the department's needs.

No other Manufacturer's software will communicate with and operate the existing PerkinElmer HPLC and components that make up the tandem PerkinElmer HPLC-ICP-MS system.

9. Identify the primary responsible staff person(s) conducting and managing this procurement. (Appropriate delegated procurement authority and completion of mandatory training required.)

*Point of contact (Place asterisk after name of person to contact for additional information).

Name	Division/Agency	Phone Number	E-mail Address
Rebecca H. Sciulli	SLD	453-5993	rebecca.sciulli@doh.hawaii.gov
Roger Hirata	DOCD / BT Branch	587-6599	roger.hirata@doh.hawaii.gov

Department shall ensure adherence to applicable administrative and statutory requirements, including HAR chapter 3-122, Subchapter 15, Cost or Pricing Data if required.

All requirements/approvals and internal controls for this expenditure is the responsibility of the department.

I certify that the information provided is to the best of my knowledge, true and correct.



Department Head Signature

2/7/14

Date

For Chief Procurement Officer Use Only

Date Notice Posted: 2/12/14

Submit written objection to this notice to issue a sole source contract within seven calendar days or as otherwise allowed from date notice posted to:

state.procurement.office@hawaii.gov

Chief Procurement Officer (CPO) Comments:

Approval is for the period 2/11/2014 – 2/12/2015 and is based on the department's representation that this equipment can only be obtain from the manufacturer PerkinElmer Health Sciences, Inc. This approval is for the solicitation process only, HRS section 103D-301(c) and HAR section 3-122-112, shall apply and award is required to be posted on the Awards Reporting System. Copies of the HCE certificate and awards posting are required to be documented in the procurement/contract file. Additionally, sole source contracts in excess of \$100,000 require cost or pricing data pursuant to HAR chapter 3-122, subchapter 15 and is required to be documented in the procurement/contract file. If there are any questions, please contact Bonnie Kahakui at 587-4702, or bonnie.a.kahakui@hawaii.gov.

Approved Disapproved No Action Required

 2/20/14
Chief Procurement Officer Signature Date