



STATE PROCUREMENT OFFICE
NOTICE & REQUEST FOR SOLE SOURCE

13 FEB -6 A8:39

STATE PROCUREMENT OFFICE
STATE OF HAWAII

TO: Chief Procurement Officer
FROM: Health State Laboratories Division
Name of Requesting Department

Pursuant to HRS §103D-306 and HAR chapter 3-122, Subchapter 9, the Department requests sole source approval to purchase the following:

1. Describe the goods, services, or construction to be procured.
VictorX Perkin Elmer (PE) multilabel reader/counter; includes Luminescence, Fluorescence Intensity (both top and bottom reading), UV-absorbance (UV-VIS) and Time-Resolved Fluorescence technologies- with Windows-compatible V ICTOR 2030 Workstation software, WorkOut 2.5 standard in Vict X4 ; temperature control standard, shaker with adjustable speed, and , installation, training, extended warranty and Quality Assurance Package.

Table with 2 columns: Vendor/Contractor/Service Provider Name, Amount of Request, Term of contract, and Prior SPO-001, Sole Source (SS) No. Includes handwritten notes like 'PerkinElmer Inc.', 'PERKINELMER HEALTH SCIENCES, INC', and dates '2/6/13'.

6. Describe in detail the following:
a. The unique features, characteristics, or capabilities of the goods, service or construction.
It is the only multilabel reader available which can provide quantitative detection using four (4) detection technologies, namely, Time-Resolved-Flourescence, , luminescence, photometry, and fluorometry using a High-Throughput format.
b. How the unique features, characteristics or capabilities of the goods, service or construction are essential for the department
It is the only system validated by the Centers for Disease Control and Prevention (CDC) Laboratory Response Network (LRN) to detect a special morphological characteristic of a BT agents and toxins from various matrices (food, clinical isolates and environmental sample).

8265

7. Describe the efforts and results in determining that this is the only vendor/contractor/service provider who can provide the

PE is phasing out the old Victors 2 and 3 models which are used for the CDC-LRN TRF assays. They have introduced a new line of instruments which are designated as the X series to replace the old models. A bridging study was performed to compare the old Victor to the new Victor X4 model. The bridging study indicates that the performance of the latter is comparable to the old model and is acceptable for use with the LRN TRF procedures. PE have indicated that they will no longer service the Victors 2 and 3 models and parts for these models will no longer be produced, thus, the purchase of an X4 is necessary. The CDC indicated that Cooperative Agreement funds may be used for the purchase of this instrument. The CDC revised the LRN procedures to incorporate the Victor X4 as an acceptable instrument and are posted in the LRN website.

8. Alternate source. Describe the other possible sources for the goods, services, or construction that were investigated but did not meet the department's needs.

There were no alternative source investigated for reasons indicated on Section 7.

9. Identify the primary responsible staff person(s) conducting and managing this procurement. (Appropriate delegated procurement authority and completion of mandatory training required.)

*Point of contact (Place asterisk after name of person to contact for additional information).

Name	Division/Agency	Phone Number	E-mail Address
Rebecca H. Sciulli	State Lab Division	808-453-5993	rebecca.sciulli@doh.hawaii.gov
			

Department shall ensure adherence to applicable administrative and statutory requirements, including HAR chapter 3-122, Subchapter 15, Cost or Pricing Data if required.

**All requirements/approvals and internal controls for this expenditure is the responsibility of the department.
I certify that the information provided is to the best of my knowledge, true and correct.**

Department Head Signature

Date

For Chief Procurement Officer Use Only

Date Notice Posted: 2/6/2013

Submit written objection to this notice to issue a sole source contract within seven calendar days or as otherwise allowed from date notice posted to:

state.procurement.office@hawaii.gov

Chief Procurement Officer (CPO) Comments:

Approval is for the period 02/06/13 to 08/30/13 and based on the department's representation that the VictorX Perkin Elmer is the only multilabel reader available providing quantitative detection and in the only system validated by the Centers for Disease Control and Prevention Laboratory Response Network. This approval is for the solicitation process only, HRS section 103D-310(c) and HAR section 3-122-112, shall apply (i.e. vendor is required to be compliant on the Hawaii Compliance Express) and award is required to be posted on the Awards Reporting System.

If there are any questions, please contact Stanton Mato at 586-0566, or stanton.d.mato@hawaii.gov.

Approved

Disapproved

No Action Required

Olamb. J. [Signature] 2/15/2013
Chief Procurement Officer Signature Date